Micro Bio Laboratory

ISO 9001 : 2015 CERTIFIED LABORATORY

.....complete microbiology

39, Asiatic Arcade, Vartak Nagar, Pokhran Road No 1, Thane(w) – 400606
Web:www.microbiolaboratory.com E–mail:microbiolab2005@yahoo.co.in, microbiolab2005@gmail.com
Mob : +91-9820384592.

ANALYTICAL REPORT

REPORT NO: 04929/04/2022

REF NO: Letter 27/04
DATE RECD : 28/04/2022

SAMPLE RECEIVED FROM: RESOLVE DIAGNOSTIC PVT LTD
ANDHERI, MUMBAI

SAMPLE NAME : BACILLUS PUMILUS
BIOLOGICAL INDICATOR

ANALYSIS REQD: STERILITY TEST

NAME OF PRIMARY PRODUCT : PLASTIC FUNNEL
(INSERTED WITH 2 BIOLOGICAL INDICATORS)

RECD . QUANTITY: 2 NOS

BATCH NO OF PRIMARY PRODUCT : ----

AGSLLP URN NO : AGS/M/22-23/0188
AGSLLP SERIAL NO OF BIOLOGICAL INDICATOR 2: RPP-16 383 & RPP-16 384
GAMMA DOSE DELIVERED : MINIMUM 25 kGy ;
IRRADIATION CYCLE DATE & TIME: 26/04/2022 AT 18:30 PM

MICROBIOLOGY TESTS [AS PER USP]

1. DESCRIPTION : TWO BIOLOGICAL INDICATORS IN A SEALED PLASTIC PACKS.

2. TEST FOR STERILITY : NO GROWTH OBSERVED IN AEROBIC MEDIA AND ANAEROBIC MEDIA.

[METHOD: DIRECT]

3. MEDIA : SOYABEAN CASEIN DIGEST MEDIUM [SCDM]
FLUID THIOGLYCOLLATE MEDIUM [FTM]

OPINION : THE SUBMITTED SAMPLE COMPLIES WITH THE TEST FOR STERILITY AS PER USP.

DATE: 14/05/2022

MicroBio Laboratory

FOR MICROBIO LABORATORY